



**CITY OF ST. LOUIS
DEPARTMENT OF FINANCE
OFFICE OF THE SUPPLY COMMISSIONER**

**CAROL L SHEPARD, CPA
SUPPLY COMMISSIONER**

**FRANCIS G. SLAY
MAYOR**

**1200 MARKET ST RM 324
ST LOUIS MO 63103
PHONE: 314-622-4580
FAX: 314-622-4141**

ADDENDUM NO. 1

65017Q0612 – GC/MS Supplies
Bid Opening Date – September 14, 2016

To: All Bidders
From: Dotlyn Bedessie, Buyer
Date: August 25, 2016
Re: 65017Q0612 – GC/MS Supplies

This addendum shall be considered part of the Specifications for bid (except as noted otherwise) and is issued to change, amplify, or delete from or otherwise explain these documents where provisions of this addendum differ from those of the original documents. This addendum shall have precedence over the original documents and shall govern.

Q1: Item #9, "Agilent HP-5MS Column or equivalent" – do you have part # or a full description of the item?

A1: The part number is 19091s-433UI through Agilent. The specifications are Agilent J&W HP-5MS, ultra inert GC column, 30m length, 0.25 mm i.d., 0.25 micrometer film, with a 7 in cage.

Q2: Would this be an "all or nothing" quote, or would we be able to bid on just a few of the items?

A2: We encourage all bids whether you are able to bid on all or just a few of the items listed.

All bidders must attach this form to the original Bid/Quote Form. This completed form will serve as acknowledgement that this information has been received.

ADDENDUM RECEIVED AND ACKNOWLEDGE

Name of Firm

Representative

Signature

Date